VALIDITY DIAGNOSTICS

APMC SYNTHETIC URINE CHALLENGE

26 OF 26 PRODUCTS FAILED

Validity Diagnostics SVT tests were performed on a Beckman Olympus AU400e in the toxicology laboratory at Advanced Pain Management Clinic, Jacksonville, FL. Samples were received as unknowns labeled 1–11 in 2020 and 1–15 in 2025.

EVERY PRODUCT TESTED FAILED

2020 · 11 ■■■■■■ ■■■■■

2025 · 15 ■■■■■■■ ■■■■■■■■

One square = one product rejected by the True Essentials Panel.

9/23/2020 — 11 SAMPLES · 11 FAILURES

	TRADITIONAL SVT			VDx TRUE ESSENTIALS			
Sample	pH	Urine Creat	SGI	Oxidant History	True Urine LD	True Urine SD	Specimen Result
Acceptable Range	3 – 10.5	> 20	> 1.003	> 10	> 10	> 10	
Perfect Urine	7.50	11.8	1.0007	0.1	3	1.5	FAIL
Dr Green Super Hero	7.30	69.6	1.013	1.6	9.3	3.4	FAIL
Quick Fix Plus	7.30	67.3	1.0125	0.1	3	4	FAIL
Ultimate Gold	6.40	39.9	1.0125	1.4	1.2	4.8	FAIL
Ultra Pure	8.70	61.6	1.0117	0.5	8.4	0.4	FAIL
Dr Greens X Agent	6.50	85.3	1.0102	2.3	9.3	4.3	FAIL
Gator Wizz	6.80	77.1	1.0105	2.1	9.3	2.1	FAIL
Golden Flask	5.50	43.8	1.014	3.3	9.3	5.9	FAIL
Monkey Flask	5.00	44.7	1.0187	0.5	5.2	4.6	FAIL
Synthetix5	8.10	35.1	1.0157	1.1	1.6	6.4	FAIL
UPASS	6.70	78.8	1.0136	2.1	9.3	8.2	FAIL

8/14/2025 — 15 SAMPLES · 15 FAILURES

	TRADITIONAL SVT			VDx TRUE ESSENTIALS			
Sample	pH	Urine Creat	SGI	Oxidant History	True Urine LD	True Urine SD	Specimen Result
Acceptable Range	3 – 10.5	> 20	> 1.003	> 10	> 10	> 10	
Test Clear PU100	7.99	98	1.0178	9.2	5.7	2.1	FAIL
QuickFix Lot 1	6.96	45	1.0185	0	6.6	5.3	FAIL
QuickLuck	6.00	31	1.0186	1.2	5.7	4.9	FAIL
Pure Gold	5.60	7.8	1.0157	0	6.6	4.8	FAIL
Xstream Lot #1	7.92	28	1.0161	0	3.9	4.1	FAIL
Pure Steam	5.59	8.7	1.0154	0	4.8	5.6	FAIL
Clean Whizz	9.38*	119	1.0125	0	3.9	5.4	FAIL
Upass	9.88*	66	1.0144	0	4.8	4.6	FAIL
Xstream Lot #2	7.01	32	1.0175	0	5.7	5.5	FAIL
Monkey Whizz	5.59	54	1.0161	1.8	56*	2.4	FAIL
Agent X	7.23	35	1.0194	0	2.1	4.1	FAIL
QuickFix Lot 2	7.27	49	1.0185	0	3.9	6.9	FAIL
Monkey Flask	9.45*	60	1.0049	0	4.8	-16	FAIL
Golden Flask	7.63	19	1.0257	0	5.7	7.1	FAIL
Sub Solution	7.56	81	1.0175	0	5.7	6.3	FAIL

Fail = specimen failed validity testing and was reported invalid. · * Possible bacterial growth but no growth visible in tubes.